



GOVERNMENT OF WEST BENGAL
Department of Health & Family Welfare

OFFICE OF THE PRINCIPAL
BANKURA SAMMILANI MEDICAL COLLEGE

Po: Kenduadhi, Dist: Bankura, W.B, PIN- 722102

Ph.No.(03242) 251324 / 244754, Fax No.(03242) 250929, E-mail: bsmc_xsa@yahoo.com

MemoNo:-BSMC/3335

Date:- 08.08.2023

Notice for Addendum (1st)

E-Tender No: BSMC/3073, Dt. 22.07.2023

e-Tender ID : 2023_HFW_552298_1

***Name of the work:* PROCUREMENT OF LAMINAR AIR FLOW HOOD**

This document should be considered by all bidders as an Addendum for the captioned tender and the following modifications and clarifications should necessarily be considered.

A. Addendum in respect of TECHNICAL SPECIFICATION: Addendum: I

- s/p -
Principal

Bankura Sammilani Medical College.
Bankura.

Memo No: BSMC/ 3335

Date: 08.08.2023

Copy forwarded for information and necessary action to the:-

1. The Director of Medical Education & Ex-Officio Secretary to the Govt. of West Bengal, Dept. of Health and Family Welfare, SwasthyaBhawan, GN-29, Sector-V, Salt Lake City, Kolkata-700 091.
2. The District Magistrate, Bankura.
3. The Medical Superintendent-cum-Vice Principal, BSMCH Bankura.
4. the HOD, Ophthalmology, BSMC, Bankura.
5. The Sub Divisional Officer, Bankura.
6. The Chairman, Bankura Municipality, Bankura
7. The Accounts Officer, BSMC, Bankura.
8. All members of the Tender & Purchase Committee, BSMCH, Bankura.
9. The Head Clerk, BSMC, Bankura
10. IT Cell, Department of Health & F.W. SwasthyaBhawan- with a request to upload this notice in the Health Departmental website.
11. The District Information Officer, NIC, Bankura, -with a request to upload this notice in the Departmental website.
12. The Managing Director, Karrot Graphics, 71-A, SP Mukherjee Road, Kol-26 for uploading the NIT in the website (www.bsmedicalcollege.org.in)
13. Notice Board.
14. Guard file.

08/08/2023
Principal

Bankura Sammilani Medical College
Bankura.

Principal
B.S. Medical College
Bankura

Standard Dimensions


Model		BST/VLAF-42	
Working Size		4'x2'x2'	
Size of HEPA Filter		4'x2'x6"	
No. of HEPA Filter		1	

[Signature]
2/05/23

Prof. & Head
Department of Ophthalmology
Bankura Sammilani Medical College
Bankura, W.B., 722102


Features:

HEPA filtration unit provides Class 100 air quality	Meet US Federal Standard 209 B (BS 5295)
Optional UV lamp to Sterilize work zone	Stainless Steel Work table
Low noise level	Made from branded parts (blower etc.)
High airflow capacity	Easy to change filters


31/05/23
Prof. & Head
Department of Ophthalmology
Bankura Sammilani Medical College
Bankura, W.B., 722102

Standard Dimensions

Cleanliness	Class 100
Velocity	90FPM \pm 20%
Hepa Filter	99.999% efficiency for particles > 0.3 μ m
Pre-Filter	85% efficiency for particles >0.5 μ m (Washable)
Particle Count	Better than US Fed Std 209B Class 10 and VDI 2083 Class3
Cabinet	Laminated High Quality Wooden Board / PCRC Sheet Powder Coated / Stainless Steel SS 304 (Optional 316 grade)
Work Table	304 Stainless Steel (Optional 316 grade)
Airflow Speed	Control Speed Controller (Three step speed Controller)
Blower	High efficient Centrifugal type with lifetime lubricated bearings
Light	High intensity, low wattage > 800 lux
Noise Level	<55dBA
Standard Accessories	Air/gas cock and mains power socket (16A)
Power Supply	220-230V, 50 Hz.
Optional Accessories	<ul style="list-style-type: none"> • Microprocessor LCD Controller • Digital Display for air flow rate • Transparent front door • Gauges Pressure (Statics Pressure Mano-Meter) • Magnahelic Gauge (for filter pressure) • U.V Germicidal tube in work area • Electronic filter choke alarm • Spare HEPA filters • Auto switch on / off for U.V Germicidal tube & Fluorescent light • Hour Meter for UV light


 24/05/23
 Prof. & Head
 Department of Ophthalmology
 Bankura Sammilani Medical College
 Bankura, W.B., 722102